

# **COVER PAGE**

**Document Type:**

Statistical Analysis Plan (SAP)

**Official study title:**

Computational Design, Fabrication, and Evaluation of Optimized Patient-Specific Transtibial Prosthetic Sockets

**NCT number:**

not yet assigned

**Document Date:**

02/15/2018

**Principal investigator:**

Prof. Hugh Herr

**Sponsor:**

NIH

**Participating sites:**

Massachusetts Institute of Technology (MIT), Cambridge, MA, USA  
McGovern Institute for Brain Research, MIT, Cambridge, MA, USA

# Statistical analysis plan

Statistical analysis of the socket clinical evaluation data will focus on the following measurements:

1. local skin contact pressures
2. metabolic power
3. gait parameters (symmetry indices for joint angles, positions, torques, and ground reaction forces)
4. the socket evaluation questionnaire

For scalar comparisons a paired t-test, with the level of significance set at 0.05, will be used to test for differences between the novel and conventional sockets. For time-series analysis of variance (ANOVA) will be used.